

SUPERVISOR TEST

NEW HARTFORD TOWN
POLICE DEPT
DATAMASTER dmt: 127606

Date: 12/19/2019

OPERATOR NAME:
MICHAEL T KOWALSKI
Permit #: 31077
Operator Agency/Dept: NEW HARTFORD TOWN PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:10
INTERNAL STANDARD	VERIFIED	14:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:10
BLANK TEST	0.00	14:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:12
BLANK TEST	0.00	14:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:13
BLANK TEST	0.00	14:14

