

SUPERVISOR TEST

NEW HARTFORD TOWN  
POLICE DEPT  
DATAMASTER dmt: 127606

Date: 04/07/2020

OPERATOR NAME:  
MICHAEL T KOWALSKI  
Permit #: 31077  
Operator Agency/Dept: NEW HARTFORD TOWN PD  
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:07
INTERNAL STANDARD	VERIFIED	19:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:08
BLANK TEST	0.00	19:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:09
BLANK TEST	0.00	19:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:10
BLANK TEST	0.00	19:11

