

SUPERVISOR TEST

NEW HARTFORD TOWN
 POLICE DEPT
 DATAMASTER dmt: 127606

Date: 02/05/2020

OPERATOR NAME:
 MICHAEL T KOWALSKI
 Permit #: 31077
 Operator Agency/Dept: NEW HARTFORD TOWN PD
 Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:11
INTERNAL STANDARD	VERIFIED	15:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:11
BLANK TEST	0.00	15:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:12
BLANK TEST	0.00	15:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:13
BLANK TEST	0.00	15:14

