

SUPERVISOR TEST

NEW HARTFORD TOWN
POLICE DEPT
DATAMASTER dmt: 127606

Date: 01/14/2020

OPERATOR NAME:
MICHAEL T KOWALSKI
Permit #: 31077
Operator Agency/Dept: NEW HARTFORD TOWN PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:04
INTERNAL STANDARD	VERIFIED	13:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.00	13:04
BLANK TEST	0.00	13:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.00	13:05
BLANK TEST	0.00	13:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.00	13:06
BLANK TEST	0.00	13:07

