

SUPERVISOR TEST

NEW HARTFORD TOWN  
POLICE DEPT  
DATAMASTER dmt: 127606

Date: 01/07/2020

OPERATOR NAME:  
MICHAEL T KOWALSKI  
Permit #: 31077  
Operator Agency/Dept: NEW HARTFORD TOWN PD  
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:50
INTERNAL STANDARD	VERIFIED	12:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.00	12:51
BLANK TEST	0.00	12:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.00	12:52
BLANK TEST	0.00	12:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.00	12:53
BLANK TEST	0.00	12:54

