

SUPERVISOR TEST

NEW YORK MILLS
POLICE DEPT.
DATAMASTER dmt: 500475

Date: 06/03/2020

OPERATOR NAME:
ROBERT C FRANKLAND
Permit #: 40674
Operator Agency/Dept: NY MILLS PD
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:51
INTERNAL STANDARD	VERIFIED	14:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:52
BLANK TEST	0.00	14:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:53
BLANK TEST	0.00	14:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:54
BLANK TEST	0.00	14:55

