

SUPERVISOR TEST

NEW YORK MILLS
POLICE DEPT.
DATAMASTER dmt: 500475

Date: 02/26/2020

OPERATOR NAME:
ROBERT C FRANKLAND
Permit #: 40674
Operator Agency/Dept: NY MILLS PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:50
INTERNAL STANDARD	VERIFIED	14:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:51
BLANK TEST	0.00	14:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:52
BLANK TEST	0.00	14:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:53
BLANK TEST	0.00	14:54

