

SUPERVISOR TEST

NEW YORK MILLS
POLICE DEPT.
DATAMASTER dmt: 500475

Date: 01/08/2020

OPERATOR NAME:
ROBERT C FRANKLAND
Permit #: 40674
Operator Agency/Dept: NY MILLS PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:22
INTERNAL STANDARD	VERIFIED	14:22
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:22
BLANK TEST	0.00	14:23
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:23
BLANK TEST	0.00	14:24
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:24
BLANK TEST	0.00	14:25

