

SUPERVISOR TEST

NEW YORK MILLS  
POLICE DEPT.  
DATAMASTER dmt: 500475

Date: 10/30/2019

OPERATOR NAME:  
ROBERT C FRANKLAND  
Permit #: 40674  
Operator Agency/Dept: NY MILLS PD  
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:35
INTERNAL STANDARD	VERIFIED	19:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	19:35
BLANK TEST	0.00	19:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	19:36
BLANK TEST	0.00	19:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	19:38
BLANK TEST	0.00	19:38

