

SUPERVISOR TEST

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 04/15/2020

OPERATOR NAME:
BRETT A GRABELDINGER
Permit #: 46896
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:43
INTERNAL STANDARD	VERIFIED	17:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:43
BLANK TEST	0.00	17:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:45
BLANK TEST	0.00	17:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:46
BLANK TEST	0.00	17:47

