

SUPERVISOR TEST

CFD

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

LATIC

Date: 05/21/2020

OPERATOR NAME:
EMRAH LATIC
Permit #: 46908
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:36
INTERNAL STANDARD	VERIFIED	00:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:37
BLANK TEST	0.00	00:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:38
BLANK TEST	0.00	00:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:39
BLANK TEST	0.00	00:40

