

SOLUTION CHANGE

SHERRILL CITY
POLICE DEPT
DATAMASTER dmt: 115108

Date: 12/04/2019

OPERATOR NAME:
ROBERT J DRAKE
Permit #: 27298
Operator Agency/Dept: SHERRILL CITY PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:32
INTERNAL STANDARD	VERIFIED	07:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:33
BLANK TEST	0.00	07:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:34
BLANK TEST	0.00	07:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:35
BLANK TEST	0.00	07:36

