

SUPERVISOR TEST

SHERRILL CITY
POLICE DEPT
DATAMASTER dmt: 115108

Date: 12/23/2019

OPERATOR NAME:
ROBERT J DRAKE
Permit #: 27298
Operator Agency/Dept: SHERRILL CITY PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:14
INTERNAL STANDARD	VERIFIED	07:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:15
BLANK TEST	0.00	07:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:16
BLANK TEST	0.00	07:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:17
BLANK TEST	0.00	07:18

