

SUPERVISOR TEST

 SHERRILL CITY
 POLICE DEPT
 DATAMASTER dmt: 115108

Date: 04/29/2020

OPERATOR NAME:
 ROBERT J DRAKE
 Permit #: 27298
 Operator Agency/Dept: SHERRILL CITY PD
 Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:14
INTERNAL STANDARD	VERIFIED	07:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:14
BLANK TEST	0.00	07:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:15
BLANK TEST	0.00	07:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:17
BLANK TEST	0.00	07:18

