

SUPERVISOR TEST

SHERRILL CITY
POLICE DEPT
DATAMASTER dmt: 115108

Date: 01/22/2020

OPERATOR NAME:
ROBERT J DRAKE
Permit #: 27298
Operator Agency/Dept: SHERRILL CITY PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:20
INTERNAL STANDARD	VERIFIED	07:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:21
BLANK TEST	0.00	07:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:22
BLANK TEST	0.00	07:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:23
BLANK TEST	0.00	07:24

