

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

Date: 06/23/2020

RMS
20-224/10

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:45
INTERNAL STANDARD	VERIFIED	14:46
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:46
BLANK TEST	0.00	14:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:47
BLANK TEST	0.00	14:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:48
BLANK TEST	0.00	14:49

