

Laser Labs, Inc.

Certificate of Accuracy Window Tint Meter and Reference Samples

This document certifies that the Tint Meter and two enclosed Reference Samples have been tested and certified to meet the manufacturers specified tolerances. The Reference Samples were given their respective stated values by a calibrated spectrophotometer at five hundred and fifty nanometers wavelength.

Date of Certification: 5/6/08

Tint Meter Model 100 Serial Number: 2051000040261

Serial Number Reference Sample 1: LO25040261

Stated Value Reference Sample 1: 20% Light Transmission

Serial Number Reference Sample 2: HI75040261

Stated Value Reference Sample 2: 77% Light Transmission

Certified by:


Jay S. Poh

Window Tint Meter Reference Samples are used to verify the accuracy of a Tint Meter. Follow the instructions to perform this test.

1. Each Reference Sample has a label affixed to it. On the label is a light transmission percentage. This percentage is referred to as the stated value of that reference sample.
2. Read the stated value.
3. Follow the test procedure for your Tint Meter and measure the light transmission of the Reference Sample.
4. The reading on the Tint Meter should not exceed two percentage points of the stated value to be within the specified accuracy ratings of a two piece Tint Meter, and should not exceed two percentage points of a one piece Tint Meter. Example: If the stated value is 25 and the meter reading was any where in between 23 to 27, then the Tint Meter has verified that it is accurate to within the specification.
5. Test your Tint Meter with each Reference Sample before any window test.
6. Keep your Reference sample protected in the Tint Meter carrying case.
7. To clean, rub with a soft 100% cotton cloth. Do not use any liquid or dry cleaners.
8. To replace, contact Laser Labs, Inc at Laser-Labs.Com or call 1-800-452-2344, for outside U.S. 508-923-6416.