## **Certificate of Accuracy**

## **Tint Meter**

This document certifies that this Tint Meter has been tested and certified to meet the manufacturer's specified tolerances. The meter was tested for accuracy and linearity as referenced to a calibrated spectrophotometer at five hundred and fifty nanometers.

Lazer Labs Quality Control

Lot Rev 5 Assembled v Calibrated Final Test i

Model 1000

Certified 01/2019 Serial # 316123733

Certified by

Edward Marcin

## **Certificate of Accuracy**

## **Tint Meter Reference Samples**

This document certifies that the enclosed tint meter reference samples have been tested and certified to meet the manufacturer's specified tolerances. The reference samples were given their respective stated values by a calibrated, NIST traceable spectrophotometer at five hundred fifty nanometers.

62%

**TEST STANDARD** 

CERTIFIED
January 2019

Serial #: 210192482 Date Tested: Laser Labs Inc.

24%

Serial #: 210192483 Date Tested: Laser Labs Inc.

Certified by

Edward Marcin